


# DETECTAMET

## Safety Data Sheet

|                       |                           |
|-----------------------|---------------------------|
| Document Reference    | 403                       |
| Date of Issue         | 12 <sup>th</sup> Feb 2025 |
| Revision Number       | 001                       |
| Date of Last Revision | 12 <sup>th</sup> Feb 2025 |

**403**

### X-Ray and Metal Detectable Plasters



#### Technical Data Sheet Applicable To:

|                     |                                                              |
|---------------------|--------------------------------------------------------------|
| <b>403-S124-X11</b> | Detectable Plasters 38mm x 38mm Pack of 100                  |
| <b>403-S731-X11</b> | Detectable Plasters 19mm x 63mm Pack of 100                  |
| <b>403-S732-X11</b> | Detectable Plasters 30mm x 64mm Pack of 100                  |
| <b>403-S733-X11</b> | Detectable Plasters 19mm x 72mm Pack of 100                  |
| <b>403-S734-X11</b> | Detectable Plasters 25mm x 72mm Pack of 100                  |
| <b>403-S735-X11</b> | Detectable Plasters 38mm x 72mm Pack of 100                  |
| <b>403-S736-X11</b> | Detectable Plasters 51mm x 72mm Pack of 100                  |
| <b>403-S737-X11</b> | Detectable Plasters Finger Pack of 50                        |
| <b>403-S738-X11</b> | Detectable Plasters Assorted Pack Multiple Sizes Pack of 100 |

**Industry Usage:**

Our range of metal & X-ray detectable plasters are blue in colour and offer a dual detection system, making them detectable by both metal and X-ray inspection systems – required under BRCGS Issue 8. They are also fully FDA 21 CER Part 807 approved.

Available in eight sizes, these detectable plasters come in packs of 50 or 100 dependent on size. We also offer an assorted pack of 100 containing multiple sizes.

**Features and Benefits:**

- Metal Detectable & X-ray Visible
- Colour: blue plasters for easy visual detection
- Shelf life: 5 years from date of production
- Sizes:
  - 38mm x 38mm (1.49 x 1.49") – Pack of 100
  - 19mm x 63mm (0.74" x 2.48") – Pack of 100
  - 30mm x 64mm (1.18 x 2.51") – Pack of 100
  - 19mm x 72mm (0.74 x 2.83") – Pack of 100
  - 25mm x 72mm (0.98 x 2.83") – Pack of 100
  - 38mm x 72mm (1.49 x 2.83") – Pack of 50
  - 51mm x 72mm (2.00 x 2.83") – Pack of 50
  - Finger 20mm x 120mm (0.78 x 4.72") – Pack of 50
  - Assorted Pack Multiple Sizes – Pack of 100
- Sizes and Quantities included in the Assorted Pack
  - 20pcs of 19 x 63mm
  - 15pcs of 19 x 72mm
  - 15pcs of 25 x 72mm
  - 10pcs of 30 x 64mm
  - 10pcs of 38 x 38mm
  - 10pcs of 38 x 72mm
  - 10pcs of 51 x 72mm
  - 10pcs of Finger 20 x 120mm

**Material and Compliance information:**

| Adhesive Type                   | Hotmalt Adhesive |
|---------------------------------|------------------|
| Coat Weight (g/m <sup>2</sup> ) | 35-45            |
| Adhesion Power (N/25mm)         | 20 +/- 3         |
| Loop Track (N/25mm)             | 15 +/- 4         |
| Shear (hrs)                     | ≥20hrs           |

**Wound Pad**

| Composition         | Gm/sqm           | 1 – non adherent layer 100% PE<br>2 – absorption layer 75% viscose 25% PP/PE<br>3 – Backing: Aluminium Sheet with a Barium Sulphate ribbon 1 mm thickness and 4 mm width |
|---------------------|------------------|--------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Total Grammage      | g/m <sup>2</sup> | 220/-10%                                                                                                                                                                 |
| Absorption Capacity | %                | ≥400                                                                                                                                                                     |
| Time of Absorption  | Sec              | ≥1                                                                                                                                                                       |

**Release Paper**

| Test      | Units            | Toll     |
|-----------|------------------|----------|
| Grammager | g/m <sup>2</sup> | 75 +/- 3 |

## Wrapping Paper

### A) Front side wrapping paper

| Test         | Units            | Toll     |
|--------------|------------------|----------|
| Total Weight | g/m <sup>2</sup> | 45 +/- 7 |

### B) Front side wrapping paper

| Test   | Units            | Toll     |
|--------|------------------|----------|
| Weight | g/m <sup>2</sup> | 44 +/- 7 |

Sealing strength for the pouches  $\geq 1.0$  N/15mm

**The validity of the certificate is subject to periodic or unexpected verification.**

Study report 2018060097 Cytotoxicity (test on extracts, DIN EN ISO 10993-5: 2009)

## Material and Methods

### Test Material

#### Spun laced non-woven/PU laminate HM + Pad Pharm pore Ultra LOT: 284/11

The test material was packed, sterilized and sent by the client. The material used as negative control was polypropylene (Lot 1717/34) and as positive control medium 6 % Dimethyl Sulfoxide (DMSO) in DMEM-FCS (Dulbecco's Modification of Eagle's Minimum Essential Medium with fetal calf serum). The reagent control was cell culture medium which was extracted at the same extract conditions as the sample.

#### Preparation of extracts (DIN EN ISO 10993-12:2012)

A surface/volume ratio of 6 cm<sup>2</sup> of test material per millilitre of extraction medium was used to perform the test. Before starting the analysis the cover paper was removed under sterile conditions. The material used for the negative control was polypropylene (0.2 g per millilitre of extraction medium), which was sterilized by autoclaving. The positive control medium was 6 % DMSO freshly prepared in DMEM-FCS before application. The estimated cytotoxicity of the DMSO solution is approximately 50 %.

**Extraction medium:** Dulbecco's Modification of Eagle's Minimum Essential Medium was used (DMEM; Fisher Scientific GmbH; Lot: 1880308). The culture medium (DMEM) was supplemented with 10 % fetal calf serum (FCS; PAN-Biotech GmbH; Lot: P122011 ). Extraction was performed with gentle shaking at  $37 \pm 1$  °C for 24,  $\pm 2$  hours.

### Preparation of cell cultures

L 929 cells (ATCC CCL 1, NCTC clone 929, connective tissue mouse, clone of strain L) were used as test cell line. The test system was chosen because this cell line has been shown to be suitable for this type of study and is recommended in the DIN EN ISO 10993-5:2009. The cells were grown in DMEM-FCS supplemented with 10 % FCS. The cells were cultured at  $37 \pm 1$  °C and 5% CO<sub>2</sub> in a humidified incubator.

24  $\pm$  2 hours before application of the extracts onto the cells, cells were harvested from standard tissue culture flasks using a trypsin/EDTA solution. The cells were resuspended in DMEM-FCS. The cell density was adjusted to  $1.5 \times 10^5$  /ml, and the wells of 96 wells tissue culture plates were inoculated with 100  $\mu$ I of cell suspension per well. The plates were incubated for 24  $\pm$  2 hours in a humidified incubator in 5 %

CO<sub>2</sub> at  $37 \pm 1$  °C. During this time the cells formed subconfluent monolayer.

### Preparation of serial dilutions of the extract

Serial dilutions of the test material extract were prepared in reagent control in order to obtain the following percentages of extract in the dilution mixture:

|         |   |      |
|---------|---|------|
| Extract | a | 100% |
|         | b | 66%  |
|         | c | 44%  |
|         | d | 30%  |
|         | e | 20%  |

Cytotoxicity testing of the test material extract was performed at all above listed levels of concentration. The extract of negative control material and the reagent control were used undiluted ("100 %") only. The positive control reagent (6 % DMSO in DMEM- FCS) was prepared in a sterile container and mixed thoroughly.

### Exposure of cells to extract and control reagents.

Each dilution of extract mentioned above and all control reagents were done as four replications. A 100  $\mu$ I aliquot of the liquids were administered into each of the four wells (using the 96 wells tissue culture plates after removing the cell culture medium from the prepared wells). As a next step, the tissue culture plates were incubated in a humidified incubator in 5 % CO<sub>2</sub> at  $37 \pm 1$  °C for at least 24 hours.

### Qualitative determination of cytotoxicity

After 24 hours of incubation, the cell culture plates were examined microscopically using phase contrast microscopy at 400 x magnification.

The degree of cytotoxicity observed in each well was numerically graded using a subjective grading system as follows:

- 0 (none reactivity):** Discrete intracytoplasmic granules; no cell lysis; no cell growth inhibition
- 1 (slight reactivity):** Not more than 20 % of the cells are round, loosely attached and without intracytoplasmic granules ore show changes in morphology; occasional lysed cells are present, only slight cell growth inhibition
- 2 (mild reactivity):** Not more than 50 % of the cells are round and devoid of intracytoplasmic granules; no extensive cell lysis and empty areas between cells; not more than 50 % cell growth inhibition
- 3 (moderate reactivity):** Not more than 70 % of cell layers contain rounded cells or are lysed; more than 50% cell growth inhibition
- 4 (severe reactivity):** Nearly complete or complete destruction of the cell layers

### Quantification of extract dilutions cytotoxicity and controls

The tetrazolium salt MTS [3-(4,5-dimethylthiazol-2-yl)-5-(3-carboxymethoxyphenyl)- 2-(4-sulfophenyl)-2H-tetrazolium] (yellow) is reduced from living cells to formazan (red brown). For this purpose, electrons are transferred from the reduction equivalents NADH (nicotinamide adenine dinucleotide) and NADPH (nicotinamide adenine dinucleotide phosphate) to MTS. NADH and NADPH are formed by mitochondrial dehydrogenases of metabolically active cells. The formazan content is thus an indirect proof for the metabolic cell activity, which increases proportionally to the cell number and should be quantified here. This method is a validated assay based on the reduction of vital dye (similar to MTT [3-(4,5-dimethylthiazol-2-yl)-2,5-diphenyltetrazolium bromide]).

After microscopically examination cell culture medium was eliminated from each well. Subsequent each well was filled with 120 µl of MTS-solution (freshly prepared). 1 blank well (without cells) of each plate was additional filled with 120 µl of MTS-solution. The plate was incubated at  $37 \pm 1$  °C and 5 % CO<sub>2</sub> for 40 min. The absorption wavelength was 490 nm and the reference wavelength was 630 nm. At first the average of blank (MV) was subtracted from each extinction measured at 490 nm and 630 nm. After this the deviation of absorption and reference wavelength was determined. The average of reagent control was estimated at 100 % cell growth. The results of other measurements were related to the reagent control. The relative inhibition of cell growth (% ICG) was calculated as follows:

$$\%ICG = 100 - 100 \times \frac{E(490 \text{ nm} - \text{MV blank}) - (630 \text{ nm} - \text{MV blank}) \text{ extract}}{E(490 \text{ nm} - \text{MV blank}) - (630 \text{ nm} - \text{MV blank}) \text{ reagent control}}$$

### Results

#### Results of the macroscopic and microscopic examinations

Cells covered with reagent control prepared did not show any damage (grade 0). Also the extract of negative control material (polypropylene) did not harm the cells

(grade 0). The positive control solution (6 % DMSO dissolved in DMEM-FCS) caused, as expected, damages (degree 3-4) of the cells.

After a 24 hours incubation of cell cultures with different concentrations of the test material extract, the following results have been observed:

#### Grade of Toxicity:

| Conc. Of extract | Microscopic examination (degree 0-4) |
|------------------|--------------------------------------|
| 100%             | 0-1                                  |
| 66%              | 0-1                                  |
| 44%              | 0                                    |
| 30%              | 0                                    |
| 20%              | 0                                    |

Slight cell growth inhibition was microscopically visible with the extract concentrations of 100% - 66%

**Measured Values of the percentage inhibition of cell growth (% ICG)**

| Controls                                                       |                 |                 |                 |       |       |       |       |       |                 |
|----------------------------------------------------------------|-----------------|-----------------|-----------------|-------|-------|-------|-------|-------|-----------------|
|                                                                | RC <sub>1</sub> | PC <sub>1</sub> | NC <sub>1</sub> | 100%  | 66%   | 44%   | 30%   | 20%   | RC <sub>1</sub> |
| <b>Measured Values</b><br>(490 nm blank) –<br>(630 nm – Blank) | 0.389           | 0.109           | 0.398           | 0.354 | 0.367 | 0.370 | 0.353 | 0.380 | 0.382           |
|                                                                | 0.370           | 0.111           | 0.391           | 0.337 | 0.341 | 0.355 | 0.405 | 0.376 | 0.437           |
|                                                                | 0.370           | 0.109           | 0.385           | 0.332 | 0.354 | 0.353 | 0.371 | 0.379 | 0.420           |
|                                                                | 0.359           | 0.108           | 0.378           | 0.334 | 0.331 | 0.353 | 0.343 | 0.377 | 0.399           |
| <b>Mean Value (MV)</b>                                         | 0.372           | 0.109           | 0.388           | 0.339 | 0.348 | 0.356 | 0.368 | 0.368 | 0.368           |
| <b>+ Standard Deviation</b>                                    | 0.012           | 0.001           | 0.009           | 0.010 | 0.016 | 0.010 | 0.027 | 0.002 | 0.024           |
| <b>ICG<sub>4</sub> in %</b>                                    | 4.8             | 72.0            | 0.7             | 13.2  | 10.9  | 8.9   | 5.8   | 3.3   | 0.0             |

<sup>1</sup> = Reagent Control (RC); <sup>2</sup> = Positive Control (PC, 6 % DMSO); <sup>3</sup> = Negative Control (NC);

<sup>4</sup> = inhibition of cell growth (ICG); \* outlier (Nalimov outlier test); Mean value of all reagent controls: **0.391**

**An inhibition of cell growth (ICG) of more than ( ) 30 % is considered as cytotoxic effect according to DIN EN ISO 10993-5.**

**Assessment of the test results**

The extract of the test material caused no relevant toxicological / biological damages to subconfluent monolayer of L929 cells under the test conditions of DIN EN ISO 10993-5:2009. The test material meets the requirements of DIN EN ISO 10993-5 and is considered non-cytotoxic.

Results of negative and positive controls confirm the sensitivity and accuracy of the test system. There was a good correlation between the colorimetric results and microscopically detectable cell appearance. The test can be considered as homogeneous and valid.

**Report 2018090909 Skin Irritation (DIN EN ISO 10993-10:2014)**

**Material and Methods**

**Test Material**

**Spunlaced non-woven HM + Pad; Pharmapore, Pharmapore IV, Cure-Aid, Cure-Aid Dressing Strips, ETO sterilized; LOT: 601/12**

The test material was sterilized, packed and sent by the client. The test procedure was performed without extraction of the test material.

## "MAGNUSSON AND KLIGMAN GRADING SCALE FOR THE EVALUATION OF CHALLENGE PATCH TEST REACTIONS"

The positive control showed the sensitization and the validity of the test system.

### Assessment of results

The test material did not cause sensitization during the course of the observation period.

### Trial


To test samples of plasters according to standard ASTM F792-08, to test the detectability of the plasters with a step wedge of different thicknesses.

### Method

Using a Stainless-Steel step wedge with of thickness 2.35, 4.25, 6.15, 8.10 and 10.5mm, the plasters are placed on top of the step wedge furthest away from the detector plate for worst case scenario. X-rayed at 100KV, 6.5MA, 60 seconds and a FFD of 1metre, HD Imaging plate scanned at 50um

### Set up

In order to create the worst possible conditions, the plasters were placed on top of the step wedged, at the remotest distance from the imaging plates, see below.



Four types of plasters were supplied, so four images were produced in the same orientation for each. Each image was identified with the thickness of the steps in the step wedge and the plaster type.

*\*No warranty is given or implied with respect to this information or patent infringement. Detectamet Ltd do not accept liability for loss or damage arising from the use of this information. Results are based on a test sample, our general experience and information from suppliers. Data and results may be confirmed by the buyer by testing for its intended conditions of use.\**

Safety You Detect

**[detectamet.global](https://detectamet.global)**

